

RA Document

| RA Details | |
|--|--|
| RA Technical End Date/Time | 24-07-2020 13:00:00 |
| RA Opening Date/Time | 24-07-2020 13:30:00 |
| RA Life Cycle (From Publish Date) | 90 (Days) |
| RA Offer Validity (From End Date) | 30 (Days) |
| Ministry/State Name | Maharashtra |
| Department Name | Rural Development Department |
| Organisation Name | Zilla Parishad Satara |
| Office Name | General Administration Dept Z P Satara 415001 |
| Total Quantity | 100 |
| Item Category | High Vacuum Suction Machine for medical purposes |
| Average Turnover (Last 3 Years) | 12 Lakh (s) |
| OEM Average Turnover (Last 3 Years) | 75 Lakh (s) |
| Experience Criteria | 3 Year (s) |
| MSE Exemption for Years of Experience and Turnover | No |
| Startup Exemption for Years of Experience and Turnover | No |
| Document required from seller | Experience Criteria,Past Performance,Bidder Turnover,OEM Authorization Certificate,OEM Annual Turnover *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer |
| Past Performance | 10 % |
| Inspection Required | No |

EMD Detail

| | |
|-------------------|-------|
| Advisory Bank | ICICI |
| EMD Percentage(%) | 0.50 |
| EMD Amount | 12425 |

ePBG Detail

| | |
|--|--|
| | |
|--|--|

| | |
|-------------------------------------|-------|
| Advisory Bank | ICICI |
| ePBG Percentage(%) | 2.00 |
| Duration of ePBG required (Months). | 2 |

(a). EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy.

(b). EMD & Performance security should be in favour of Beneficiary, wherever it is applicable.

Beneficiary:

District Health Officer ZP Satara
 General Administration Dept Z P Satara 415001, Rural Development Department, Zilla Parishad Satara,
 (Aniruddha Achyut Athalye)

Splitting

RA splitting not applied.

1. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU / Public Listed Company for number of years as indicated in the bid document before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the year. In case of bunch bids, the category of primary product having highest value should meet this criterion.
2. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.
3. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 10% of bid quantity, in at least one of the last three years before the bid opening date to any Central / State Govt Organization / PSU / Public Listed Company. Copies of relevant contracts (proving supply of cumulative order quantity in any one year) to be submitted along with bid in support of quantity supplied in the relevant year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

High Vacuum Suction Machine For Medical Purposes (100 pieces)

Technical Specifications

* As per GeM Category Specification

| Specification | Specification Name | Values | Bid Requirement (Allowed Values) |
|-------------------------|--|---|----------------------------------|
| GENERAL FEATURES | Product description | High Vacuum suction machine -Medical (Aspiration unit) | * |
| | Clinical purpose | Suction may be used to clear the respiratory tract/airway of blood, saliva, vomit, or other secretions so that a patient may breathe, also to remove the blood from the area where surgery is to be carried out | * |
| | Application | Minor OT, Surgery, Intensive care unit, Emergency department, Wards, Obstetrics, Gastroenterology, Plastic surgery | * |
| MATERIAL & CONSTRUCTION | Material of cabinet (Base, top and panel) | ABS Plastic moulded | ABS Plastic moulded |
| | Rust proof housing cabinet and castor wheels | Yes | * |
| | The unit is mounted in such a way as to produce minimum vibrations | Yes | * |
| | Material of twin jars/bottles | unbreakable Polysulfone (PS) | unbreakable Polysulfone (PS) |
| | Suction bottles/jars with Synthetic rubber lid and fitted with arrangement to prevent overflow of fluid to outside | Yes | * |
| | Capacity of each suction bottle/jar graduated (ml) | 4000 | 4000 |
| | Interconnected jar/bottle with overflow protection device | Yes | * |
| Double overflow | Yes | * | |

| | | | |
|--------------------------------------|--|------------------------------------|------------------------------------|
| | protection with safety device provided | | |
| | Non-collapsible suction tubing ID (mm) | 8 | 8 |
| | Material of suction tube | Silicon | Silicon |
| | Minimum Length of suction tube (mtr) | 2 | * |
| | Flutter free vacuum control knob provided | Yes | * |
| | Bacterial suction filter/air filters and water traps between the suction bottle and suction pump provided | Yes | * |
| | Material of tray | Stainless Steel (SS) | * |
| | Vacuum Gauge (Bourdon tube type) with concentric scale marking in black on white background, mounted in panel, easily visible during use of machine | "10Cm dia, 0 to 760mmHg" | "10Cm dia, 0 to 760mmHg" |
| | Entire unit mounted on four castor wheels with minimum two breaks | Yes | * |
| | Material of castor wheels | Thermoplastic Rubber(TPR) | Thermoplastic Rubber(TPR) |
| | Facility for easy maintenance | Back open | * |
| | Size (HxWxL) (mmxmmxmm) | - | * |
| | Weight (Kg) | 20 to 30 | * |
| OPERATIONAL & TECHNICAL REQUIREMENTS | Input power supply | 220+/-10% V, 50Hz, Single phase AC | 220+/-10% V, 50Hz, Single phase AC |
| | Power consumption (Watt) | 380 | * |
| | Power indicator | Yes | * |
| | Integrated on/off switch on the trolley | Yes | * |
| | In-built maintenance free battery of suitable capacity provided | No | No |
| | Minimum battery backup on full | NA | NA |

| | | | |
|------------------------------|---|---|---------------------------|
| | charge (minutes) | | |
| | Power cable for usage in ambulance/car provided | No | No |
| | Motor capacity (HP) | 0.5 | 0.5 |
| | The motor shall conform to IS:996/1979 | Yes | * |
| | Type of suction pump | Oil free | Oil free |
| | Heavy duty motorized pump for fast vacuum built-up | Yes | * |
| | Minimum air flow displacement of suction pump (Ltrs/Minute) | 25 to 30 | 25 to 30 |
| | Built-up Vacuum (Maximum) | "-700 +/-10mmHg (-93kPa)" | "-700 +/-10mmHg (-93kPa)" |
| | Sound level (Maximum) (dB) | 55 | * |
| STANDARD ACCESSORIES | Compatible 3 core power cord with Indian plug | 2 meter length | 2 meter length |
| | Jars/ Bottles (No) | 2 | * |
| | Suction tubing (set) | 1 | 1 |
| PACKING MODE | The product is packed individually in a wooden/corrugated box including all the accessories in such a way that there will be no transit damage during transportation till destination | Yes | * |
| ENVIRONMENTAL CONSIDERATIONS | Storage conditions | Temperature 0 deg C to 50 deg C, RH 10% to 90% | * |
| | Work or usage environment | Temperature 10 deg C to 50 deg C, RH 10% to 90% | * |
| MARKING | Product is ISI Marked | No | No |
| | BIS License No (CM/L No) | No | * |
| CERTIFICATIONS & REPORTS | Submission of Test Report from OEM complying the declared | YES | * |

| | | | |
|------------------------|---|---|----------------------|
| | specification at the time of supply | | |
| | Product certification | BIS,EU-CE,US-FDA | * |
| | Four digit number of notified body If product is EU-CE certified | - | * |
| | Certificate No | - | * |
| | Certificate Date | - | * |
| | Certificate issuing authority | - | * |
| | Certification, performance and safety standards specific to the device | IEC 60601-1:2012,IEC 6060-1-2:2014,IS:4533/1995,IS:7080 (Part-3)/1992 | * |
| | Submission of all the certifications and test reports to the buyer along with supplies on demand | Yes | * |
| WARRANTY & MAINTENANCE | Onsite Warranty (Year) | 2 | 2, 3, 4, 5 Or higher |
| | User technical and maintenance manual detailing complete maintaining schedule with routine maintenance should be provided | Yes | * |
| | Contact details of manufacturer, supplier and local service agent to be provided | Yes | * |

* Specifications highlighted in bold are the Golden Parameters.

* Bidders may note that In respect of non-golden Parameters, the specifications 'Values' chosen by Buyer will generally be preferred over 'Bid requirement (allowed Values) by the Buyer.

Additional Specification Documents

Consignee/Reporting Officer and Quantity

| S.No. | Consignee/Reporting Officer | Address | Quantity | Delivery Days | |
|-------|-----------------------------|--|----------|---------------|--|
| 1 | Aniruddha Achyut Athalye | 415001,Zilla Parishad Satara, Satara-Koregaon Road,Sadar Bazar Camp.Satara | 100 | 15 | |

Bid Specific Additional Terms and Conditions

1. Bidders can also submit the EMD with Payment online through RTGS / internet banking in Beneficiary name District Integrated Society for Health & FW Program Satara Account No. 646001701438 IFSC Code ICIC0006460 Bank Name ICICI Satara Branch address Ch Shahu Stadium Branch Satara.
Bidder to indicate bid number and name of bidding entity in the transaction details field at the time of on-line transfer. Bidder has to upload scanned copy / proof of the Online Payment Transfer along with bid.
2. Dedicated /toll Free Telephone No. for Service Support : BIDDER/OEM must have Dedicated/toll Free Telephone No. for Service Support.
3. Availability of Service Centres: Bidder/OEM must have a Functional Service Centre in the State of each Consignee's Location in case of carry-in warranty. (Not applicable in case of goods having on-site warranty). If service center is not already there at the time of bidding, successful bidder / OEM shall have to establish one within 30 days of award of contract. Payment shall be released only after submission of documentary evidence of having Functional Service Centre.
4. Scope of supply (Bid price to include all cost components) : Supply Installation Testing Commissioning of Goods and Training of operators and providing Statutory Clearances required (if any)
5. The Bidder / OEM {themselves or through reseller(s)}, should have executed project for supply and installation / commissioning of same or similar Category Products during preceding 3 financial years (i.e. current year and three previous financial years) as on opening of bid, as per following criteria:
 - (i) Single order of at least 35% of estimated bid value; or
 - (ii) Two orders of at least 20% each of estimated bid value; or
 - (iii) Three orders of at least 15% each of estimated bid value.Satisfactory Performance certificate issued by respective Buyer Organization for the above Orders should be uploaded with bid. In case of bunch bids, the Category related to primary product having highest bid value should meet this criterion

[This RA is also governed by the General Terms and Conditions](#)

---Thank You---