

RA Document

RA Details	
RA Technical End Date/Time	27-07-2020 14:00:00
RA Opening Date/Time	27-07-2020 14:30:00
RA Life Cycle (From Publish Date)	90 (Days)
RA Offer Validity (From End Date)	30 (Days)
Ministry/State Name	Maharashtra
Department Name	Rural Development Department
Organisation Name	Zilla Parishad Satara
Office Name	General Administration Dept Z P Satara 415001
Total Quantity	100
Item Category	Infusion Pump Volumetric/ Syringe Based
Average Turnover (Last 3 Years)	40 Lakh (s)
OEM Average Turnover (Last 3 Years)	200 Lakh (s)
Experience Criteria	3 Year (s)
MSE Exemption for Years of Experience and Turnover	No
Startup Exemption for Years of Experience and Turnover	No
Document required from seller	Experience Criteria,Past Performance,Bidder Turnover,Certificate (Requested in ATC),OEM Authorization Certificate,OEM Annual Turnover *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
Past Performance	10 %
Inspection Required	No

EMD Detail

Advisory Bank	ICICI
EMD Percentage(%)	0.50
EMD Amount	41440

ePBG Detail

--	--

Advisory Bank	ICICI
ePBG Percentage(%)	2.00
Duration of ePBG required (Months).	12

(a). EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy.

(b). EMD & Performance security should be in favour of Beneficiary, wherever it is applicable.

Beneficiary:

District Health Officer ZP Satara
 General Administration Dept Z P Satara 415001, Rural Development Department, Zilla Parishad Satara,
 (Aniruddha Achyut Athalye)

Splitting

RA splitting not applied.

1. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU / Public Listed Company for number of years as indicated in the bid document before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the year. In case of bunch bids, the category of primary product having highest value should meet this criterion.
2. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.
3. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 10% of bid quantity, in at least one of the last three years before the bid opening date to any Central / State Govt Organization / PSU / Public Listed Company. Copies of relevant contracts (proving supply of cumulative order quantity in any one year) to be submitted along with bid in support of quantity supplied in the relevant year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

Infusion Pump Volumetric/ Syringe Based (100 pieces)

Technical Specifications

* As per GeM Category Specification

Specification	Specification Name	Values	Bid Requirement (Allowed Values)
Standards	Availability of any of Certification such as ISO / European-CE / US-FDA / CE	YES	YES
	Details of the Certifications such as type of certification, number, date and validity	-	*
	Conformity to Standards	YES	*
Performance Parameters	Type of Infusion Pump	Syringe Infusion Pump	Syringe Infusion Pump
	Technology (of operation)	Programmable Mode	Programmable Mode
	Flow rate Range with increment in ml/hr	0.1-99.9 ml/hr(0.1 ml/hr increment) for Volumetric Infusion Pump Only	0.1-99.9 ml/hr(0.1 ml/hr increment) for Volumetric Infusion Pump Only
	Infused Volume Range in ml	0.1-999(0.1 ml/hr increment)	0.1-999(0.1 ml/hr increment)
	Flow rate & volume	flow rate & volume to be infused should be configured to limit the maximum allowable range	*
	KVO (Keep Vein Open) Rate in ml/hr	3	3
	Accuracy in %	2	2
	Bolus Type	Bolus with volume pre-selection	Bolus with volume pre-selection
	Bolus Rate Range in ml/hr	800 ml/hr for 50 ml syringe	800 ml/hr for 50 ml syringe
	Bolus Volume in ml	0.1 ml to 5 ml in 0.1 ml increments(auto bolus)	0.1 ml to 5 ml in 0.1 ml increments(auto bolus)
	Display Type	LED	LED
	Size of Display (length x width) (mm x mm)	-	*
	Contents of Display	-	*
Alphanumeric Programming Key Board	Yes	*	

Operating and Transportation Temperature range	-	*
Humidity Conditions (operating and transport & storage) (15 to 90%)	YES	*
Pressure Conditions(operating and transport & storage) (50 to 106 kpa)	YES	*
Power sources	Battery operated	*
Battery Type	Rechargeable lithium polymer	Rechargeable lithium polymer
Battery Rating	-	*
Battery run time at normal delivery rate in hours	8	*
Battery run time at highest delivery rate in hours	2-3 hrs	*
Battery recharge time (full charge) in hours	4	*
Power Supply 220-240 V AC, 50 Hz) (3 wire grounded system)	YES	*
Electrical Protection (Class I, Type CF)	YES	*
Fuse Type	220V 2A*2, 12V 2A*2	220V 2A*2, 12V 2A*2
Maximum Power Consumption in watt	20	*
Fitting of device on the IV bottle and bag	No	No
IR Sensor to detect the abnormality in IV Bag or bottle	No	No
Audible and Visual alarm facility	Should have an audible and visual alarm with automatic pump stop function for the following : pressure drop, air standby, volume infused, door open, battery Pre-alarm, recall, invalid rate, empty etc	*
Availability of Infusion mode	Yes	Yes

	setting(rate mode,drip mode,wait mode,time mode)		
	Protection against fluid ingress - IP*1 - Protected against vertically falling drops of water	YES	YES
	Capability of Motioning of empty Intravenous bags or bottles	No	No
	Warranty in Years	1	1
Dimensional & Material Parameters	Dimension (Excluding power cables)(mm x mm x mm)	-	*
	Weight(with clamp base)in kg	2.3	*
	Material (Outer Shell)	ABS PLASTIC	*
Additional Parameters	Pole clamp Multi-function mounting clamp	YES	*
	Quick titration of rate or dose with volume timing programming	YES	*
	Compatible with any standard (PVC) infusion sets available in local indian market	YES	*
	Features :	(a) memory retention (b) air sensor (c) history records (d) night inner lighting function (e)pressure dynamic display (f) wireless communication function (g) dual cpu monitoring function	*
	No of infusion sets supplied with the single unit	1	*
	Packing (Individual)	YES	*
Reports	Type tests reports from a Central Govt lab OR International Laboratory Accreditation Cooperation (ILAC)or	YES	*

	their worldwide affiliated/ recognized labs OR NABL approved lab		
	Test Report No	-	*
	Test Report Date	-	*
	Name and complete address of the Lab	-	*
	Copy of certificates to be provided to buyer on demand after placement of order (Applicable if availability of test report indicated as YES, otherwise put NA)	YES	*

* Specifications highlighted in bold are the Golden Parameters.

* Bidders may note that In respect of non-golden Parameters, the specifications 'Values' chosen by Buyer will generally be preferred over 'Bid requirement (allowed Values) by the Buyer.

Additional Specification Documents

Consignee/Reporting Officer and Quantity

S.No.	Consignee/Reporting Officer	Address	Quantity	Delivery Days	
1	Aniruddha Achyut Athalye	415001,Zilla Parishad Satara, Satara-Koregaon Road,Sadar Bazar Camp.Satara	100	15	

Bid Specific Additional Terms and Conditions

1. IMPORTED PRODUCTS: In case of imported products, OEM or Authorized Seller of OEM should have a registered office in India to provide after sales service support in India. The certificate to this effect should be submitted.
2. Bidders can also submit the EMD with Payment online through RTGS / internet banking in Beneficiary name District Integrated Society for Health & FW Program Satara Account No. 646001701438 IFSC Code ICIC0006460 Bank Name ICICI Satara Branch address Ch Shahu Stadium Branch Satara.
Bidder to indicate bid number and name of bidding entity in the transaction details field at the time of on-line transfer. Bidder has to upload scanned copy / proof of the Online Payment Transfer along with bid.
3. Timely Servicing / rectification of defects during warranty period: After having been notified of the defects / service requirement during warranty period, Seller has to complete the required Service / Rectification within 15 days time limit. If the Seller fails to complete service / rectification with defined time limit, a penalty of 0.5% of Unit Price of the product shall be charged as penalty for each week of delay from the seller. Seller can deposit the penalty with the Buyer directly else the Buyer shall have a right to recover all such penalty amount from the Performance Security (PBG).Cumulative Penalty cannot exceed more than 10% of the total contract value after which the Buyer shall have the right to get the service /

rectification done from alternate sources at the risk and cost of the Seller besides forfeiture of PBG. Seller shall be liable to re-imburse the cost of such service / rectification to the Buyer.

4. Bidder / OEM has to give an undertaking that after expiry of warranty period, it will provide AMC Service for next 1 years for the offered products at the rate not more than 1 % of contract price per annum. Buyer reserves the right to enter into an AMC agreement (covering preventive maintenance and servicing)with the Successful Bidder / OEM after expiry of the Warranty period at rate as mentioned above and the payment for the AMC charges would be made Annually after rendering of the AMC Services of the relevant AMC period. Performance Security of the successful bidder shall be forfeited if it fails to accept the AMC contract when called upon by the buyer. The original Performance Security of contract will be returned only after submission and verification of AMC Performance Security for 2% of total AMC value valid up to AMC period plus 2 months (if there is no other claim). (Undertaking of acceptance to be uploaded with bid).
5. Availability of Service Centres: Bidder/OEM must have a Functional Service Centre in the State of each Consignee's Location in case of carry-in warranty. (Not applicable in case of goods having on-site warranty). If service center is not already there at the time of bidding, successful bidder / OEM shall have to establish one within 30 days of award of contract. Payment shall be released only after submission of documentary evidence of having Functional Service Centre.
6. The Bidder / OEM {themselves or through reseller(s)}, should have executed project for supply and installation / commissioning of same or similar Category Products during preceding 3 financial years (i.e. current year and three previous financial years) as on opening of bid, as per following criteria:

- (i) Single order of at least 35% of estimated bid value; or
- (ii) Two orders of at least 20% each of estimated bid value; or
- (iii) Three orders of at least 15% each of estimated bid value.

Satisfactory Performance certificate issued by respective Buyer Organization for the above Orders should be uploaded with bid. In case of bunch bids, the Category related to primary product having highest bid value should meet this criterion

[This RA is also governed by the General Terms and Conditions](#)

---Thank You---