

RA Document

RA Details	
RA Technical End Date/Time	31-07-2020 18:00:00
RA Opening Date/Time	31-07-2020 18:30:00
RA Life Cycle (From Publish Date)	90 (Days)
RA Offer Validity (From End Date)	30 (Days)
Ministry/State Name	Maharashtra
Department Name	Rural Development Department
Organisation Name	Zilla Parishad Satara
Office Name	General Administration Dept Z P Satara 415001
Total Quantity	800
Item Category	Patient Bed (Fowler and Semi Fowler),patient care mattresses
Average Turnover (Last 3 Years)	35 Lakh (s)
OEM Average Turnover (Last 3 Years)	200 Lakh (s)
Experience Criteria	3 Year (s)
MSE Exemption for Years of Experience and Turnover	No
Startup Exemption for Years of Experience and Turnover	No
Document required from seller	Experience Criteria,Past Performance,Bidder Turnover,OEM Authorization Certificate,OEM Annual Turnover *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
Primary Product	Patient Bed (Fowler and Semi Fowler)
Past Performance	10 %
Inspection Required	No

EMD Detail

Advisory Bank	ICICI
EMD Percentage(%)	1.00
EMD Amount	73996

ePBG Detail

Advisory Bank	ICICI
ePBG Percentage(%)	2.00
Duration of ePBG required (Months).	2

(a). EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy.

(b). EMD & Performance security should be in favour of Beneficiary, wherever it is applicable.

Beneficiary:

District Health Officer ZP Satara
General Administration Dept Z P Satara 415001, Rural Development Department, Zilla Parishad Satara,
(Aniruddha Achyut Athalye)

Splitting

RA splitting not applied.

1. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU / Public Listed Company for number of years as indicated in the bid document before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the year. In case of bunch bids, the category of primary product having highest value should meet this criterion.

2. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

3. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 10% of bid quantity, in at least one of the last three years before the bid opening date to any Central / State Govt Organization / PSU / Public Listed Company. Copies of relevant contracts (proving supply of cumulative order quantity in any one year) to be submitted along with bid in support of quantity supplied in the relevant year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

Patient Bed (Fowler And Semi Fowler) (400 pieces)

Technical Specifications

[* As per GeM Category Specification](#)

Specification	Specification Name	Values	Bid Requirement (Allowed Values)
Performance Parameters	Type of bed	Semifowler	Semifowler
	Number of bed platform sections	2 for semi fowler bed	2 for semi fowler bed
	Availability of Backrest section	Yes	*
	Backrest section linkage	welded with MS CRCA tube (14 G)	*
	Availability of Leg section	NA for semifowler	NA for semifowler
	Leg section linkage	NA for semifowler	NA for semifowler
	Mattress platform shall be perforated	Yes	*
	Screw mechanism	welded with ERW MS tube (18 G) in M.S. cover made from (16 G) ERW tube	*
	Manual adjustments	Backrest through two screw systems with thrust bearings individually manouvred by handle for semifowler bed	Backrest through two screw systems with thrust bearings individually manouvred by handle for semifowler bed
	Back rest - min 45 % of the frame length	Yes	*
	The bottom end of the H legs are provided with PVC shoes	Yes in case of bed without castors	Yes in case of bed without castors
	The bed has head & foot panels detachable by hand without need of any tool	Yes	*
	Saline rod provision	four locations on the bed platform to hold stainless steel Telescopic Saline rod 12mm dia with 15.87 mm dia x 1.2 mm (18 G) stainless steel	*
Provision for mounting of syringe pump	outer covering tube with a knob to mount	*	

	syringe pump		
Finishing & workmanship shall be of high standard	Yes	*	
All corners shall be rounded off so that there shall be no sharp corners and holes should be burr free	Yes	*	
Coating for metal parts	M.S. tubular parts, linkages, flats are to be In-house, pretreated / shot blasted and Epoxy powder coated with coating thickness 50 to 100 microns	*	
Material of handle	Plastic	Plastic	
Number of detachable moulded PP with PVC insert head and foot side panels	4	*	
Provision of castors	NA (No castors provided)	NA (No castors provided)	
Number of hooks in Telescopic I V Pole	4	4	
Urine Bag Holder	Yes	*	
Moulded Chart Holder	Yes	*	
Provision of mattress	No	*	
Number of sections in mattress	NA (incase mattress is not provided)	NA (incase mattress is not provided)	
Provision of collapsible Railing	Yes	Yes	
Material and Dimensional Parameters	Overall Length of bed in mm with 5% tolerance	2050 to 2100	*
	Overall width of bed in mm with 5% tolerance	900 to 950	*
	Size of Buffer in mm	75	75
	Bed Platform frame size with 5% tolerance	2070 mm L x 910 mm W	*
	Mattress Platform size with 5% tolerance	1970 mm L x 790 mm W	*
	Buffer to Handle Length with 5% tolerance	2275 mm	*
	Fixed Height	600 mm	*
Material and size of section	1.2 mm (18 G) CRCA M.S perforated sheet	*	

	top	
Overall length of Back Rest	775-800	775-800
Overall width of back rest	750-775	*
Overall length of Knee Rest	NA for semifowler	NA for semifowler
Overall width of Knee Rest	NA for semifowler	NA for semifowler
Overall length of Leg Rest	550-600	*
Overall width of Leg Rest	NA for semifowler	NA for semifowler
Overall length of Fix Section	100-150	100-150
Overall width of fix section	750-775	750-775
Raised Backrest Angle	65°	*
Raised Kneerest Angle	NA for semifowler	*
Bed frame tube size	60 mm x 30 mm x 1.6 mm (16 G) Thick ERW tube	*
Dimensions for Fitting of frame	ERW tubes diameter 31.75 mm x 1.2 mm (18 G) thick and diameter 22.22 mm x 1.2 mm (18 G) thick and MS sheet having thickness 3.0 mm (11 G)	*
Diameter of Four corner rubber buffers in mm	75	*
Material of head and foot panels	SS Tubular	*
Diameter of MS tubular in mm	30 to 50 mm	*
Diameter of SS Laminated	30 to 50 mm	*
Patient Working Load	135 Kg	*
Safe Working Load	150 Kg	*
Thickness of PU foam of mattress in inch	NA (incase mattress is not provided)	*
Density of PU foam of mattress in Kg/m ³	NA (incase mattress is not provided)	*
Material of mattress	NA (incase mattress is	*

	cover of mattress	not provided)	
	Colour of mattress cover	Blue	Blue
	Material of collapsing railing	NA (incase railing not provided)	NA (incase railing not provided)
	Types of collapsing railing	NA (incase railing not provided)	NA (incase railing not provided)
Reports	Manufacturer Certificates	ISO 9001:2015	*
	Details of manufacturer certificates such as number and date	-	*
	Availability of certificates of quality system	ISO 13485	*
	Availability of test report from govt/ NABL/ILAC accredited lab covering all parameters	No	No
	Test report number , Incase test reports not available , put NA	-	*
	Test report date, Incase test reports not available , put NA	-	*
	Name of test lab , Incase test reports not available , put NA	-	*
	Address of test lab , Incase test reports not available , put NA	-	*
	Copies of all certifications and reports to be provided to buyer on demand at time of supplies	NA ,Incase test reports not available	NA ,Incase test reports not available

* Specifications highlighted in bold are the Golden Parameters.

* Bidders may note that In respect of non-golden Parameters, the specifications 'Values' chosen by Buyer will generally be preferred over 'Bid requirement (allowed Values) by the Buyer.

Additional Specification Documents

Consignee/Reporting Officer and Quantity

S.No.	Consignee/Reporti	Address	Quantity	Delivery Days
-------	-------------------	---------	----------	---------------

	ng Officer			
1	Aniruddha Achyut Athalye	415001,Zilla Parishad Satara, Satara-Koregaon Road,Sadar Bazar Camp.Satara	400	15

Patient Care Mattresses (400 pieces)

Technical Specifications

[* As per GeM Category Specification](#)

Specification	Specification Name	Values	Bid Requirement (Allowed Values)
Features	Material of Mattress	poly urethane	*
	Mattress with Sections	-	*
	Fire Retardan,Anti Microbial and Water proof (as per IPX - 54 standard) properties of cover for Mattress	Yes	Yes
	Color of Cover Material	Grey	*
	Cover Material for Mattress	-	*
	Density of Material of Mattress in (Kg/m3)	32	32, 40
	Dimensions	Overall Length (mm, Feet)	1980 mm(6.5 feet)
Overall Width (mm)		762 mm(2.5 feet)	762 mm(2.5 feet), 915 mm(3 feet)
Thickness of Mattress (mm)		75 mm +/-0.5%	75 mm +/-0.5%
Product details, Certifications	Product Code available on Web	No	*
	Product Code No.	-	*
	Drawing No. available on Web	No	*
	Drawing No.	-	*
	Availability Of Test Certificate From Govt Lab/NABL Or ILAC Accredited Lab For Firwe Retardant And Microbial Properties Of Cover Material Of Mattress	No	*

Test certificate No. and Date	-	*
Name and Address of the Lab. where tested	-	*

* Specifications highlighted in bold are the Golden Parameters.

* Bidders may note that In respect of non-golden Parameters, the specifications 'Values' chosen by Buyer will generally be preferred over 'Bid requirement (allowed Values) by the Buyer.

Additional Specification Documents

Consignee/Reporting Officer and Quantity

S.No.	Consignee/Reporting Officer	Address	Quantity	Delivery Days	
1	Aniruddha Achyut Athalye	415001,Zilla Parishad Satara, Satara-Koregaon Road,Sadar Bazar Camp.Satara	400	15	

Bid Specific Additional Terms and Conditions

1. After award of contract – Successful Bidder shall have to get advance sample approved from buyer before bulk manufacturing / starting bulk supplies. Successful Bidder shall submit 5 samples for Buyer's approval, within 5 days of award of contract. Buyer shall, as per contract specifications framework, either approve the advance sample or will provide complete list of modification required in the sample within 5 days of receipt of advance sample. Seller shall be required to ensure supply as per approved sample with modifications as communicated by Buyer. If there is delay from buyer side in approval of advance sample – the delivery period shall be refixed without LD for the period of delay in sample approval. In case, the sample is found to have major deviations / not conforming to the Contract specifications, the buyer at its discretion may call for fresh samples for approval before allowing bulk supplies or may terminate the contract after notifying the deviations to the seller. Unless otherwise provided in the contract, all samples required for test shall be supplied by the contractor free of cost. Where under the contract, the contractor is required to submit an advance sample, any expenses incurred by the contractor on or in connection with the production of stores in bulk, before the sample has been approved unconditionally shall be borne by the Seller and he shall not claim any compensation in the event of such sample being found unacceptable by the Buyer / Consignee.
2. Scope of supply (Bid price to include all cost components) : Supply Installation Testing and Commissioning of Goods
3. Bidder Turn Over Criteria: The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.
4. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be

uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria. In case of bunch bids, the OEM of CATEGORY RELATED TO primary product having highest bid value should meet this criterion.

5. Bidder / OEM has to give an undertaking that after expiry of warranty period, it will provide AMC Service for next 1 years for the offered products at the rate not more than 1 % of contract price per annum. Buyer reserves the right to enter into an AMC agreement (covering preventive maintenance and servicing)with the Successful Bidder / OEM after expiry of the Warranty period at rate as mentioned above and the payment for the AMC charges would be made Annually after rendering of the AMC Services of the relevant AMC period. Performance Security of the successful bidder shall be forfeited if it fails to accept the AMC contract when called upon by the buyer. The original Performance Security of contract will be returned only after submission and verification of AMC Performance Security for 2% of total AMC value valid up to AMC period plus 2 months (if there is no other claim). (Undertaking of acceptance to be uploaded with bid).
6. Bidders can also submit the EMD with Payment online through RTGS / internet banking in Beneficiary name District Integrated Society for Health & FW Program Satara Account No. 646001701438 IFSC Code ICIC0006460 Bank Name ICICI Branch address Ch Shahu Stadium Branch Satara.
Bidder to indicate bid number and name of bidding entity in the transaction details field at the time of on-line transfer. Bidder has to upload scanned copy / proof of the Online Payment Transfer along with bid.

[This RA is also governed by the General Terms and Conditions](#)

---Thank You---